## NOTES:

SUBSTRATE

CORNING: FUSED SILICA 458/678

2. ROHS COMPLIANT

3. WAVELENGTH RANGE: 350 - 2200nm

4. COATING (APPLY ACROSS COATING APERTURE)


S1 & S2: UV-VIS

 $R(AVG) \le 1.5\%$  FROM 250-700nm @ 0° AOI  $R(ABS) \le 1.0\%$  FROM 350-450nm @ 0° AOI

## 5. FINE GRIND SURFACE

6. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE

7. IMAGE ORIENTATION: BEAM DEVIATION



## FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

	S1	\$2	
SHAPE	PLANO	PLANO	
SURFACE QUALITY	20 - 10	20 - 10	
MIN CLEAR APERTURE	Ø11.25	Ø11.25	
MIN COATING APERTURE	N/A	N/A	
POWER AT 632.8nm	0.5 RINGS	0.5 RINGS	
IRREGULARITY AT 632.8nm	0.2 RING	0.2 RING	
BEVEL	PROTECTIVE, AS NEEDED	PROTECTIVE, AS NEEDED	

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

		<b>Edmund Optics</b> ®		
THIRD ANG PROJECTIO		TITLE	PRISM WEDGE FS 5 DEG 12.5MM UV VIS	
ALL DIMS IN	mm	DWG NO	35855	SHEET 1 OF 1